



Benannt durch/Designated by
Zentralstelle der Länder
für Gesundheitsschutz
bei Arzneimitteln und
Medizinprodukten
www.zlg.de
BS-MDR-099



EU Quality Management System Certificate

Regulation (EU) 2017/745 on Medical Devices, Annex IX Chapter I

Certificate No. G15 044841 0031 Rev. 00

Manufacturer:

Interacoustics A/S

Audiometer Allé 1
5500 Middelfart
DENMARK

SRN Manufacturer - DK-MF-000001216

The quality management system has been evaluated in accordance with Regulation (EU) 2017/745, Annex IX Chapter I with a positive result.

Details on devices covered by the quality management system are described on the following page(s). The report referenced below summarises the results of the assessment and includes reference to relevant CS, harmonised standards and test reports.

The certified quality management system is subject to periodical surveillance.

If class I devices in sterile conditions, with measuring function, or reusable surgical instruments are covered by this certificate, the audit was limited to the respective aspects relating to

- establishing, securing, and maintaining sterile conditions,
- conformity of the devices with the metrological requirements,
- reuse of the device, in particular cleaning, disinfection, sterilization, maintenance and functional testing and the related instructions for use.

If class IIa or class IIb devices are covered by this certificate, the quality management system assessment was accompanied by the assessment of technical documentation for devices selected on a representative basis. The periodical surveillance includes further assessment of the technical documentation on the basis of representative samples.

If class III (excluding custom-made implantable devices) or class IIb implantable devices are covered by this certificate, an EU Technical Documentation Assessment Certificate in accordance with Annex IX Chapter II is required before placing them on the market.

All applicable requirements of the Testing, Certification, Validation and Verification Regulations TÜV SÜD Group have to be complied with.

For details and certificate validity see: www.tuvsud.com/ps-cert?q=cert:G15 044841 0031 Rev. 00

Report No.: 713408723

Preceding Certificate No.: G10 044841 0027 Rev. 00

Valid from: 2026-09-10

Valid until: 2031-09-09

Christoph Dicks

Head of Certification/Notified Body

Issue date: 2026-08-24



Benannt durch/Designated by
 Zentralstelle der Länder
 für Gesundheitsschutz
 bei Arzneimitteln und
 Medizinprodukten
 www.zlg.de
 BS-MDR-099



EU Quality Management System Certificate

Regulation (EU) 2017/745 on Medical Devices, Annex IX Chapter I

Certificate No. G15 044841 0031 Rev. 00

Classification: Class IIa
Device Group: MDA 0204 - Other active non-implantable devices for monitoring and/or diagnosis

Classification: Class IIa
Device Group: MDA 0315 - Software

The validity of this certificate depends on conditions and/or is limited to the following: - none -

Revision History:

Rev.	Dated	Report	Description
00	2026-09-10	713408723	Renewal of certificate Administrative merge / transfer to new Certificate Type